

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041423\
 Data File : VN077344.D
 Acq On : 14 Apr 2023 11:14
 Operator : JC\MD
 Sample : VN0414WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0414WBL01

Quant Time: Apr 17 05:07:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N040723W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Apr 07 13:04:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.825	128	48685	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.107	114	345136	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	355094	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.583	65	116081	28.114	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	93.700%
60) 4-Bromofluorobenzene	12.854	95	164100	26.726	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	89.100%
63) Toluene-d8	10.566	98	376177	30.135	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.433%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041423\
Data File : VN077344.D
Acq On : 14 Apr 2023 11:14
Operator : JC\MD
Sample : VN0414WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0414WBL01

Quant Time: Apr 17 05:07:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N040723W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Apr 07 13:04:49 2023
Response via : Initial Calibration

