

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041423\
 Data File : VN077345.D
 Acq On : 14 Apr 2023 11:52
 Operator : JC\MD
 Sample : 02325-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 230412032-07-TRIP-BLANK

Quant Time: Apr 17 05:08:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N040723W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Apr 07 13:04:49 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 04/17/2023
 Supervised By : Mahesh Dadoda 04/17/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.819	128	47453	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.107	114	358619	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	361191	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.583	65	119124	29.600	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	98.667%
60) 4-Bromofluorobenzene	12.854	95	159376	25.518	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	85.067%
63) Toluene-d8	10.572	98	393611	30.999	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	103.333%
Target Compounds						
						Qvalue
15) Acetone	4.443	58	7628m	8.257	ug/l	
25) Chloroform	7.978	83	10272m	0.975	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041423\
Data File : VN077345.D
Acq On : 14 Apr 2023 11:52
Operator : JC\MD
Sample : 02325-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
230412032-07-TRIP-BLANK

Quant Time: Apr 17 05:08:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N040723W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Apr 07 13:04:49 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/17/2023
Supervised By :Mahesh Dadoda 04/17/2023

