


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041423\
Data File : VN077340.D
Acq On : 14 Apr 2023 08:56
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N040723W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Fri Apr 07 13:04:49 2023



AutoFind: Scans 1855, 1856, 1857; Background Corrected with Scan 1847

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.2	21376	PASS
75	95	30	60	52.3	61568	PASS
95	95	100	100	100.0	117747	PASS
96	95	5	9	7.0	8256	PASS
173	174	0.00	2	0.7	583	PASS
174	95	50	100	74.1	87280	PASS
175	174	5	9	6.6	5788	PASS
176	174	95	101	96.5	84224	PASS
177	176	5	9	6.4	5429	PASS