

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041425\
 Data File : VN086272.D
 Acq On : 14 Apr 2025 16:26
 Operator : JC\MD
 Sample : PB167573ZHE#12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167573ZHE#12

Manual Integrations
 APPROVED

Reviewed By :John Carlone 04/16/2025
 Supervised By :Semsettin Yesilyurt 04/16/2025

Quant Time: Apr 15 03:13:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N041125W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Apr 12 01:21:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.812	128	33474	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	195825	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	174832	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.577	65	99822	30.574	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery = 101.900%			
60) 4-Bromofluorobenzene	12.847	95	97028	27.875	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery = 92.900%			
63) Toluene-d8	10.565	98	294656	29.853	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery = 99.500%			
Target Compounds						
						Qvalue
12) Methyl Acetate	5.024	43	4576	1.271	ug/l	100
15) Acetone	4.430	58	12172	31.592	ug/l	89
18) Methylene Chloride	5.271	84	8524	2.933	ug/l #	87
44) Isopropyl Acetate	8.683	43	45887m	6.309	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041425\
Data File : VN086272.D
Acq On : 14 Apr 2025 16:26
Operator : JC\MD
Sample : PB167573ZHE#12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167573ZHE#12

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/16/2025
Supervised By :Semsettin Yesilyurt 04/16/2025

Quant Time: Apr 15 03:13:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N041125W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Apr 12 01:21:52 2025
Response via : Initial Calibration

