

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041519\
 Data File : VN055048.D
 Acq On : 15 Apr 2019 17:48
 Operator : JC/SP
 Sample : VN0415WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0415WBL01

Quant Time: Apr 16 06:41:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 12:28:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	547010	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	865273	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	711037	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	215465	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	340791	46.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.74%	
35) Dibromofluoromethane	7.59	113	275628	47.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.32%	
50) Toluene-d8	10.09	98	1021151	49.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.68%	
62) 4-Bromofluorobenzene	12.40	95	294761	43.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.62%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041519\
Data File : VN055048.D
Acq On : 15 Apr 2019 17:48
Operator : JC/SP
Sample : VN0415WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0415WBL01

Quant Time: Apr 16 06:41:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 12:28:28 2019
Response via : Initial Calibration

