

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041519\
 Data File : VN055056.D
 Acq On : 15 Apr 2019 21:20
 Operator : JC/SP
 Sample : K2199-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-02-041219-A

Quant Time: Apr 16 07:04:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 12:28:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	531220	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	842492	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	712411	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	235756	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	334194	47.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.66%	
35) Dibromofluoromethane	7.59	113	267034	47.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.86%	
50) Toluene-d8	10.09	98	1023212	50.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.56%	
62) 4-Bromofluorobenzene	12.40	95	298817	45.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.22%	
Target Compounds						
16) Acetone	3.82	43	14048	7.330	ug/l	98
20) Methylene Chloride	4.54	84	48989	7.271	ug/l	96
43) Isopropyl Acetate	8.17	43	14044	1.410	ug/l	96
95) Naphthalene	15.13	128	10807	4.279	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041519\
Data File : VN055056.D
Acq On : 15 Apr 2019 21:20
Operator : JC/SP
Sample : K2199-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-02-041219-A

Quant Time: Apr 16 07:04:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 12:28:28 2019
Response via : Initial Calibration

