

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041519\
 Data File : VN055057.D
 Acq On : 15 Apr 2019 21:46
 Operator : JC/SP
 Sample : K2199-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-02-041219-B

Quant Time: Apr 16 07:08:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 12:28:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	516595	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	828053	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	699675	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	232122	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	336634	49.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.04%	
35) Dibromofluoromethane	7.59	113	263220	47.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.12%	
50) Toluene-d8	10.09	98	991332	50.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.10%	
62) 4-Bromofluorobenzene	12.40	95	290382	45.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.20%	
Target Compounds						
16) Acetone	3.82	43	14941	8.017	ug/l	96
20) Methylene Chloride	4.54	84	425551	64.952	ug/l	96
43) Isopropyl Acetate	8.17	43	14953	1.527	ug/l	99
95) Naphthalene	15.13	128	75324	11.567	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041519\
Data File : VN055057.D
Acq On : 15 Apr 2019 21:46
Operator : JC/SP
Sample : K2199-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-02-041219-B

Quant Time: Apr 16 07:08:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 12:28:28 2019
Response via : Initial Calibration

