

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041519\
 Data File : VN055061.D
 Acq On : 15 Apr 2019 23:32
 Operator : JC/SP
 Sample : K2390-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 200050

Quant Time: Apr 16 07:24:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 12:28:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	532954	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	850598	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	711603	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	223599	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	332413	46.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.84%	
35) Dibromofluoromethane	7.59	113	270854	47.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.30%	
50) Toluene-d8	10.09	98	1019711	50.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.24%	
62) 4-Bromofluorobenzene	12.40	95	296261	44.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.58%	
Target Compounds						
16) Acetone	3.81	43	12337	6.416	ug/l	92
20) Methylene Chloride	4.55	84	163804	24.234	ug/l	97
43) Isopropyl Acetate	8.17	43	12476	1.240	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041519\
Data File : VN055061.D
Acq On : 15 Apr 2019 23:32
Operator : JC/SP
Sample : K2390-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
200050

Quant Time: Apr 16 07:24:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 12:28:28 2019
Response via : Initial Calibration

