

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041519\
 Data File : VN055064.D
 Acq On : 16 Apr 2019 00:51
 Operator : JC/SP
 Sample : K2407-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-17-S

Quant Time: Apr 16 07:33:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 12:28:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	527359	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	831070	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	715553	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	239535	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	332566	47.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.88%	
35) Dibromofluoromethane	7.59	113	267035	48.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.16%	
50) Toluene-d8	10.09	98	1005830	50.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.20%	
62) 4-Bromofluorobenzene	12.40	95	300319	46.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.94%	
Target Compounds						
20) Methylene Chloride	4.54	84	443884	66.367	ug/l	96
43) Isopropyl Acetate	8.17	43	16088	1.637	ug/l #	95
95) Naphthalene	15.13	128	10808	4.260	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041519\
Data File : VN055064.D
Acq On : 16 Apr 2019 00:51
Operator : JC/SP
Sample : K2407-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-17-S

Quant Time: Apr 16 07:33:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 12:28:28 2019
Response via : Initial Calibration

