

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN041519\
 Data File : VN055047.D
 Acq On : 15 Apr 2019 16:51
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 16 06:09:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 12:28:28 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.568	0.535	5.8	99	0.00
3 P	Chloromethane	0.741	0.666	10.1	96	0.00
4 C	Vinyl Chloride	0.623	0.562	9.8#	98	0.00
5 T	Bromomethane	0.352	0.309	12.2	95	0.00
6 T	Chloroethane	0.336	0.303	9.8	97	0.00
7 T	Trichlorofluoromethane	0.779	0.720	7.6	99	0.00
8 T	Diethyl Ether	0.249	0.226	9.2	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.408	0.383	6.1	102	0.00
10 T	Methyl Iodide	0.633	0.582	8.1	102	0.00
11 T	Tert butyl alcohol	0.036	0.034	5.6	97	0.00
12 CM	1,1-Dichloroethene	0.397	0.357	10.1#	98	0.00
13 T	Acrolein	0.045	0.036	20.0	90	0.00
14 T	Allyl chloride	0.971	0.909	6.4	96	0.00
15 T	Acrylonitrile	0.218	0.203	6.9	93	0.00
16 T	Acetone	0.180	0.175	2.8	108	0.00
17 T	Carbon Disulfide	1.281	1.146	10.5	95	0.00
18 T	Methyl Acetate	0.579	0.482	16.8	95	0.00
19 T	Methyl tert-butyl Ether	1.420	1.336	5.9	96	0.00
20 T	Methylene Chloride	0.634	0.544	14.2	97	0.00
21 T	trans-1,2-Dichloroethene	0.537	0.488	9.1	97	0.00
22 T	Diisopropyl ether	1.997	1.901	4.8	97	0.00
23 T	Vinyl Acetate	1.320	1.330	-0.8	95	0.00
24 P	1,1-Dichloroethane	1.063	0.995	6.4	97	0.00
25 T	2-Butanone	0.269	0.268	0.4	97	0.00
26 T	2,2-Dichloropropane	0.671	0.673	-0.3	102	0.00
27 T	cis-1,2-Dichloroethene	0.607	0.562	7.4	95	0.00
28 T	Bromochloromethane	0.513	0.468	8.8	98	0.00
29 T	Tetrahydrofuran	0.177	0.165	6.8	93	0.00
30 C	Chloroform	1.033	0.942	8.8#	97	0.00
31 T	Cyclohexane	1.158	0.932	19.5	100	0.00
32 T	1,1,1-Trichloroethane	0.818	0.763	6.7	97	0.00
33 S	1,2-Dichloroethane-d4	0.665	0.608	8.6	97	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.334	0.321	3.9	97	0.00
36 T	1,1-Dichloropropene	0.471	0.467	0.8	99	0.00
37 T	Ethyl Acetate	0.384	0.362	5.7	95	0.00
38 T	Carbon Tetrachloride	0.448	0.414	7.6	96	0.00
39 T	Methylcyclohexane	0.511	0.518	-1.4	102	0.00
40 TM	Benzene	1.421	1.396	1.8	97	0.00
41 T	Methacrylonitrile	0.213	0.211	0.9	98	0.00
42 TM	1,2-Dichloroethane	0.493	0.478	3.0	97	0.00
43 T	Isopropyl Acetate	0.591	0.597	-1.0	94	0.00
44 TM	Trichloroethene	0.356	0.355	0.3	99	0.00
45 C	1,2-Dichloropropane	0.396	0.387	2.3#	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN041519\
 Data File : VN055047.D
 Acq On : 15 Apr 2019 16:51
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 16 06:09:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 12:28:28 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.226	0.218	3.5	94	0.00
47 T	Bromodichloromethane	0.465	0.460	1.1	97	0.00
48 T	Methyl methacrylate	0.293	0.298	-1.7	97	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	111	0.00
50 S	Toluene-d8	1.196	1.189	0.6	97	0.00
51 T	4-Methyl-2-Pentanone	0.342	0.347	-1.5	94	0.00
52 CM	Toluene	0.814	0.833	-2.3#	99	0.00
53 T	t-1,3-Dichloropropene	0.406	0.429	-5.7	96	0.00
54 T	cis-1,3-Dichloropropene	0.504	0.534	-6.0	98	0.00
55 T	1,1,2-Trichloroethane	0.313	0.308	1.6	97	0.00
56 T	Ethyl methacrylate	0.376	0.395	-5.1	93	0.00
57 T	1,3-Dichloropropane	0.534	0.542	-1.5	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.139	0.150	-7.9	94	0.00
59 T	2-Hexanone	0.215	0.225	-4.7	97	0.00
60 T	Dibromochloromethane	0.306	0.314	-2.6	95	0.00
61 T	1,2-Dibromoethane	0.289	0.293	-1.4	96	0.00
62 S	4-Bromofluorobenzene	0.389	0.406	-4.4	99	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
64 T	Tetrachloroethene	0.407	0.388	4.7	102	0.00
65 PM	Chlorobenzene	1.006	0.980	2.6	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.370	0.360	2.7	96	0.00
67 C	Ethyl Benzene	1.797	1.806	-0.5#	100	0.00
68 T	m/p-Xylenes	0.651	0.672	-3.2	100	0.00
69 T	o-Xylene	0.630	0.650	-3.2	98	0.00
70 T	Styrene	1.020	1.093	-7.2	98	0.00
71 P	Bromoform	0.212	0.225	-6.1	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	4.437	3.910	11.9	100	0.00
74 T	N-amyl acetate	1.249	1.216	2.6	96	0.00
75 P	1,1,2,2-Tetrachloroethane	1.187	0.987	16.8	95	0.00
76 T	1,2,3-Trichloropropane	0.899	0.853	5.1	106	0.00
77 T	Bromobenzene	1.106	0.951	14.0	97	0.00
78 T	n-propylbenzene	4.794	4.432	7.6	101	0.00
79 T	2-Chlorotoluene	2.990	2.662	11.0	100	0.00
80 T	1,3,5-Trimethylbenzene	3.581	3.231	9.8	100	0.00
81 T	trans-1,4-Dichloro-2-butene	0.225	0.216	4.0	96	0.00
82 T	4-Chlorotoluene	2.790	2.631	5.7	101	0.00
83 T	tert-Butylbenzene	3.024	2.691	11.0	100	0.00
84 T	1,2,4-Trimethylbenzene	3.393	3.219	5.1	101	0.00
85 T	sec-Butylbenzene	3.853	3.536	8.2	101	0.00
86 T	p-Isopropyltoluene	3.306	3.139	5.1	102	0.00
87 T	1,3-Dichlorobenzene	1.681	1.622	3.5	100	0.00
88 T	1,4-Dichlorobenzene	1.603	1.563	2.5	100	0.00
89 T	n-Butylbenzene	2.454	2.598	-5.9	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN041519\
Data File : VN055047.D
Acq On : 15 Apr 2019 16:51
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Apr 16 06:09:11 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 12:28:28 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.538	0.478	11.2	99	0.00
91 T	1,2-Dichlorobenzene	1.692	1.598	5.6	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.146	0.137	6.2	94	0.00
93 T	1,2,4-Trichlorobenzene	0.644	0.723	-12.3	101	0.00
94 T	Hexachlorobutadiene	0.522	0.509	2.5	107	0.00
95 T	Naphthalene	1.432	1.549	-8.2	94	0.00
96 T	1,2,3-Trichlorobenzene	0.653	0.722	-10.6	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6