

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN041519\
 Data File : VN055066.D
 Acq On : 16 Apr 2019 1:44
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 16 07:40:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 12:28:28 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	90	0.00
2 T	Dichlorodifluoromethane	0.568	0.559	1.6	91	0.00
3 P	Chloromethane	0.741	0.734	0.9	94	0.00
4 C	Vinyl Chloride	0.623	0.605	2.9#	94	0.00
5 T	Bromomethane	0.352	0.332	5.7	90	0.00
6 T	Chloroethane	0.336	0.335	0.3	94	0.00
7 T	Trichlorofluoromethane	0.779	0.751	3.6	91	0.00
8 T	Diethyl Ether	0.249	0.243	2.4	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.408	0.382	6.4	90	0.00
10 T	Methyl Iodide	0.633	0.595	6.0	92	0.00
11 T	Tert butyl alcohol	0.036	0.031	13.9	78	0.00
12 CM	1,1-Dichloroethene	0.397	0.378	4.8#	91	0.00
13 T	Acrolein	0.045	0.041	8.9	90	0.00
14 T	Allyl chloride	0.971	0.966	0.5	90	0.00
15 T	Acrylonitrile	0.218	0.219	-0.5	89	0.00
16 T	Acetone	0.180	0.135	25.0	74	0.00
17 T	Carbon Disulfide	1.281	1.195	6.7	88	0.00
18 T	Methyl Acetate	0.579	0.530	8.5	93	0.00
19 T	Methyl tert-butyl Ether	1.420	1.440	-1.4	91	0.00
20 T	Methylene Chloride	0.634	0.601	5.2	95	0.00
21 T	trans-1,2-Dichloroethene	0.537	0.533	0.7	93	0.00
22 T	Diisopropyl ether	1.997	2.068	-3.6	94	0.00
23 T	Vinyl Acetate	1.320	1.446	-9.5	92	0.00
24 P	1,1-Dichloroethane	1.063	1.075	-1.1	93	0.00
25 T	2-Butanone	0.269	0.258	4.1	83	0.00
26 T	2,2-Dichloropropane	0.671	0.605	9.8	81	0.00
27 T	cis-1,2-Dichloroethene	0.607	0.614	-1.2	92	0.00
28 T	Bromochloromethane	0.513	0.509	0.8	95	0.00
29 T	Tetrahydrofuran	0.177	0.174	1.7	86	0.00
30 C	Chloroform	1.033	1.035	-0.2#	94	0.00
31 T	Cyclohexane	1.158	0.973	16.0	93	0.00
32 T	1,1,1-Trichloroethane	0.818	0.842	-2.9	95	0.00
33 S	1,2-Dichloroethane-d4	0.665	0.675	-1.5	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.334	0.326	2.4	95	0.00
36 T	1,1-Dichloropropene	0.471	0.446	5.3	92	0.00
37 T	Ethyl Acetate	0.384	0.347	9.6	88	0.00
38 T	Carbon Tetrachloride	0.448	0.417	6.9	93	0.00
39 T	Methylcyclohexane	0.511	0.473	7.4	90	0.00
40 TM	Benzene	1.421	1.386	2.5	93	0.00
41 T	Methacrylonitrile	0.213	0.222	-4.2	99	0.00
42 TM	1,2-Dichloroethane	0.493	0.478	3.0	93	0.00
43 T	Isopropyl Acetate	0.591	0.590	0.2	90	0.00
44 TM	Trichloroethene	0.356	0.352	1.1	94	0.00
45 C	1,2-Dichloropropane	0.396	0.387	2.3#	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN041519\
 Data File : VN055066.D
 Acq On : 16 Apr 2019 1:44
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 16 07:40:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 12:28:28 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.226	0.218	3.5	91	0.00
47 T	Bromodichloromethane	0.465	0.458	1.5	93	0.00
48 T	Methyl methacrylate	0.293	0.300	-2.4	94	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	85	0.00
50 S	Toluene-d8	1.196	1.203	-0.6	95	0.00
51 T	4-Methyl-2-Pentanone	0.342	0.340	0.6	89	0.00
52 CM	Toluene	0.814	0.824	-1.2#	94	0.00
53 T	t-1,3-Dichloropropene	0.406	0.427	-5.2	92	0.00
54 T	cis-1,3-Dichloropropene	0.504	0.531	-5.4	94	0.00
55 T	1,1,2-Trichloroethane	0.313	0.310	1.0	94	0.00
56 T	Ethyl methacrylate	0.376	0.410	-9.0	93	0.00
57 T	1,3-Dichloropropane	0.534	0.538	-0.7	93	0.00
58 T	2-Chloroethyl Vinyl ether	0.139	0.155	-11.5	93	0.00
59 T	2-Hexanone	0.215	0.211	1.9	88	0.00
60 T	Dibromochloromethane	0.306	0.316	-3.3	92	0.00
61 T	1,2-Dibromoethane	0.289	0.293	-1.4	93	0.00
62 S	4-Bromofluorobenzene	0.389	0.419	-7.7	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64 T	Tetrachloroethene	0.407	0.372	8.6	95	0.00
65 PM	Chlorobenzene	1.006	0.976	3.0	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.370	0.366	1.1	95	0.00
67 C	Ethyl Benzene	1.797	1.768	1.6#	95	0.00
68 T	m/p-Xylenes	0.651	0.658	-1.1	94	0.00
69 T	o-Xylene	0.630	0.647	-2.7	95	0.00
70 T	Styrene	1.020	1.096	-7.5	96	0.00
71 P	Bromoform	0.212	0.225	-6.1	92	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
73 T	Isopropylbenzene	4.437	3.790	14.6	95	0.00
74 T	N-amyl acetate	1.249	1.235	1.1	95	0.00
75 P	1,1,2,2-Tetrachloroethane	1.187	0.978	17.6	91	0.00
76 T	1,2,3-Trichloropropane	0.899	0.790	12.1	96	0.00
77 T	Bromobenzene	1.106	0.966	12.7	96	0.00
78 T	n-propylbenzene	4.794	4.257	11.2	95	0.00
79 T	2-Chlorotoluene	2.990	2.642	11.6	96	0.00
80 T	1,3,5-Trimethylbenzene	3.581	3.174	11.4	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.225	0.198	12.0	86	0.00
82 T	4-Chlorotoluene	2.790	2.572	7.8	96	0.00
83 T	tert-Butylbenzene	3.024	2.683	11.3	97	0.00
84 T	1,2,4-Trimethylbenzene	3.393	3.170	6.6	97	0.00
85 T	sec-Butylbenzene	3.853	3.431	11.0	96	0.00
86 T	p-Isopropyltoluene	3.306	3.041	8.0	96	0.00
87 T	1,3-Dichlorobenzene	1.681	1.598	4.9	96	0.00
88 T	1,4-Dichlorobenzene	1.603	1.540	3.9	96	0.00
89 T	n-Butylbenzene	2.454	2.457	-0.1	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN041519\
Data File : VN055066.D
Acq On : 16 Apr 2019 1:44
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Apr 16 07:40:04 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 12:28:28 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.538	0.483	10.2	98	0.00
91 T	1,2-Dichlorobenzene	1.692	1.568	7.3	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.146	0.131	10.3	88	0.00
93 T	1,2,4-Trichlorobenzene	0.644	0.708	-9.9	96	0.00
94 T	Hexachlorobutadiene	0.522	0.490	6.1	100	0.00
95 T	Naphthalene	1.432	1.571	-9.7	93	0.00
96 T	1,2,3-Trichlorobenzene	0.653	0.716	-9.6	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6