

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041519\
 Data File : VN055044.D
 Acq On : 15 Apr 2019 15:05
 Operator : JC/SP
 Sample : K2356-02
 Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-1A

Quant Time: Apr 16 06:02:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 12:28:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	500844	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	790937	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	651991	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	247316	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.02	65	311883	46.85	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	93.70%	
35) Dibromofluoromethane	7.59	113	249006	47.11	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	94.22%	
50) Toluene-d8	10.09	98	953461	50.40	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	100.80%	
62) 4-Bromofluorobenzene	12.40	95	281970	45.85	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	91.70%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041519\
Data File : VN055044.D
Acq On : 15 Apr 2019 15:05
Operator : JC/SP
Sample : K2356-02
Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-1A

Quant Time: Apr 16 06:02:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 12:28:28 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.66 min Scan# 1831

Delta R.T. -0.00 min

Lab File: VN055044.D

Acq: 15 Apr 2019 15:05

Instrument :

MSVOA_N

ClientSampled :


S-1A

Tot Ion:168 Resp: 500844

Ion Ratio Lower Upper

168 100

99 54.4 41.9 62.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.66

200000



150000

100000

50000

0

Time-->



#33

1,2-Dichloroethane-d4

Concen: 46.847 ug/l

RT: 8.02 min Scan# 1944

Delta R.T. -0.00 min

Lab File: VN055044.D


Acq: 15 Apr 2019 15:05

Tgt Ion: 65 Resp: 311883

Ion Ratio Lower Upper

65 100

67 52.4 0.0 106.4



Abundance Ion 64.90 (64.60 to 65.60): VNO

Ion 66.90 (66.60 to 67.60): VNO

8.02


150000


100000

50000

0

Time-->





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.58 min Scan# 2118

Delta R.T. 0.00 min

Lab File: VN055044.D

Acq: 15 Apr 2019 15:05

Instrument :

MSVOA_N

ClientSampleId :

S-1A


Tot Ion:114 Resp: 790937

Ion Ratio Lower Upper

114 100

63 22.2 0.0 45.2




88 15.8 0.0 31.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN0

Ion 87.90 (87.60 to 88.60): VN0



#35

Dibromofluoromethane

Concen: 47.108 ug/l

RT: 7.59 min Scan# 1808

Delta R.T. -0.00 min

Lab File: VN055044.D

Acq: 15 Apr 2019 15:05


Tgt Ion:113 Resp: 249006

Ion Ratio Lower Upper

113 100

111 102.9 82.1 123.1



192 21.0 16.5 24.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 50.398 ug/l

RT: 10.09 min Scan# 2586

Delta R.T. 0.00 min

Lab File: VN055044.D

Acq: 15 Apr 2019 15:05

Instrument :

MSVOA_N

ClientSampled :


S-1A

Tot Ion: 98 Resp: 953461

Ion Ratio Lower Upper

98 100

100 63.2 51.1 76.7



Abundance Ion 98.00 (97.70 to 98.70): VN055038.D (-2569) (-)

Ion 100.00 (99.70 to 100.70): VN055038.D (-2569) (-)

10.09

500000

400000



300000

200000

100000

0

Time-->



#62

4-Bromofluorobenzene

Concen: 45.846 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. 0.00 min

Lab File: VN055044.D

Acq: 15 Apr 2019 15:05


Tgt Ion: 95 Resp: 281970

Ion Ratio Lower Upper

95 100

174 85.1 0.0 174.4

176 82.6 0.0 166.8



Abundance Ion 94.90 (94.60 to 95.60): VN055038.D (-3291) (-)

Ion 173.80 (173.50 to 174.50): VN055038.D (-3291) (-)

Ion 175.80 (175.50 to 176.50): VN055038.D (-3291) (-)

12.40

200000


150000


100000

50000

0

Time-->









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. 0.00 min
Lab File: VN055044.D
Acq: 15 Apr 2019 15:05

Instrument :
MSVOA_N
ClientSampled :
S-1A

Tot Ion:117 Resp: 651991
Ion Ratio Lower Upper
117 100
82 56.9 44.6 66.8
119 31.7 25.5 38.3




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.34 min Scan# 3598
Delta R.T. 0.00 min
Lab File: VN055044.D
Acq: 15 Apr 2019 15:05

Tgt Ion:152 Resp: 247316
Ion Ratio Lower Upper
152 100
115 57.1 28.5 85.6
150 156.7 0.0 347.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

