

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041520\
 Data File : VN060954.D
 Acq On : 15 Apr 2020 12:28
 Operator : JC/MD
 Sample : PB128175TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB128175TB

Quant Time: Apr 16 07:11:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	130625	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	221564	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	207040	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	91023	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	85976	47.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.90%	
35) Dibromofluoromethane	7.56	113	65658	49.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.02%	
50) Toluene-d8	10.08	98	273549	49.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.98%	
62) 4-Bromofluorobenzene	12.40	95	100606	50.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.32%	
Target Compounds						
16) Acetone	3.78	43	11668	18.370	ug/l	97
20) Methylene Chloride	4.53	84	2112	1.352	ug/l	91
43) Isopropyl Acetate	8.14	43	49667	12.991	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041520\
Data File : VN060954.D
Acq On : 15 Apr 2020 12:28
Operator : JC/MD
Sample : PB128175TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB128175TB

Quant Time: Apr 16 07:11:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

