

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041520\
 Data File : VN060957.D
 Acq On : 15 Apr 2020 13:36
 Operator : JC/MD
 Sample : L2230-66
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-25-31-39

Manual Integrations
 APPROVED

MMDadoda
 4/16/2020 11:15:42 AM

Quant Time: Apr 16 07:17:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	152432	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	258815	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	242225	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	107249	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	101652	48.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.16%	
35) Dibromofluoromethane	7.56	113	76502	48.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.76%	
50) Toluene-d8	10.08	98	320621	50.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.32%	
62) 4-Bromofluorobenzene	12.40	95	120507	51.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.88%	
Target Compounds						
16) Acetone	3.78	43	24163	32.599	ug/l	100
20) Methylene Chloride	4.51	84	3419m	1.875	ug/l	
25) 2-Butanone	6.80	43	10318	8.602	ug/l	93
43) Isopropyl Acetate	8.14	43	41638	9.324	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041520\
Data File : VN060957.D
Acq On : 15 Apr 2020 13:36
Operator : JC/MD
Sample : L2230-66
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-25-31-39

Manual Integrations
APPROVED

MMDadoda
4/16/2020 11:15:42 AM

Quant Time: Apr 16 07:17:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

