

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041520\
 Data File : VN060959.D
 Acq On : 15 Apr 2020 14:21
 Operator : JC/MD
 Sample : L2122-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-04-041020

Quant Time: Apr 16 07:22:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	143407	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	244066	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	226818	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	99873	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	93769	47.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.26%	
35) Dibromofluoromethane	7.56	113	72621	49.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.42%	
50) Toluene-d8	10.08	98	300353	49.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.66%	
62) 4-Bromofluorobenzene	12.40	95	112176	50.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.56%	
Target Compounds						
16) Acetone	3.78	43	12389	17.766	ug/l	99
20) Methylene Chloride	4.53	84	2875	1.676	ug/l #	93
43) Isopropyl Acetate	8.14	43	43332	10.289	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041520\
Data File : VN060959.D
Acq On : 15 Apr 2020 14:21
Operator : JC/MD
Sample : L2122-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-04-041020

Quant Time: Apr 16 07:22:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

