

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041520\
 Data File : VN060963.D
 Acq On : 15 Apr 2020 15:52
 Operator : JC/MD
 Sample : L2245-13
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-23B

Quant Time: Apr 16 07:31:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	134857	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	227920	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	218243	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	97479	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	90890	49.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.20%	
35) Dibromofluoromethane	7.56	113	68278	49.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.08%	
50) Toluene-d8	10.08	98	288195	51.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.40%	
62) 4-Bromofluorobenzene	12.39	95	108045	52.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.74%	
Target Compounds						
16) Acetone	3.78	43	24083	36.725	ug/l	97
20) Methylene Chloride	4.52	84	2819	1.748	ug/l	87
25) 2-Butanone	6.81	43	7139	6.728	ug/l	96
43) Isopropyl Acetate	8.14	43	47992	12.203	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041520\
Data File : VN060963.D
Acq On : 15 Apr 2020 15:52
Operator : JC/MD
Sample : L2245-13
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-23B

Quant Time: Apr 16 07:31:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

