

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041520\
 Data File : VN060969.D
 Acq On : 15 Apr 2020 18:08
 Operator : JC/MD
 Sample : L2245-37
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-27

Quant Time: Apr 16 07:43:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	142082	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	243320	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	231121	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	101495	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	95782	49.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.22%	
35) Dibromofluoromethane	7.56	113	72901	49.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.10%	
50) Toluene-d8	10.08	98	304646	50.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.38%	
62) 4-Bromofluorobenzene	12.40	95	113334	51.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.92%	
Target Compounds						
16) Acetone	3.78	43	13731	19.874	ug/l	99
20) Methylene Chloride	4.53	84	1921	1.130	ug/l	89
43) Isopropyl Acetate	8.14	43	47951	11.421	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041520\
Data File : VN060969.D
Acq On : 15 Apr 2020 18:08
Operator : JC/MD
Sample : L2245-37
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-27

Quant Time: Apr 16 07:43:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

