

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041520\
 Data File : VN060971.D
 Acq On : 15 Apr 2020 18:53
 Operator : JC/MD
 Sample : L2230-54
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1-9-17

Quant Time: Apr 16 07:48:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	137203	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	232861	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	220489	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	97629	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	93118	49.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.88%	
35) Dibromofluoromethane	7.56	113	69797	49.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.14%	
50) Toluene-d8	10.08	98	292266	50.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.64%	
62) 4-Bromofluorobenzene	12.40	95	108784	51.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.22%	
Target Compounds						
16) Acetone	3.78	43	21122	31.659	ug/l	96
20) Methylene Chloride	4.52	84	2703	1.647	ug/l	88
25) 2-Butanone	6.80	43	6771	6.272	ug/l	99
43) Isopropyl Acetate	8.14	43	41972	10.446	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041520\
Data File : VN060971.D
Acq On : 15 Apr 2020 18:53
Operator : JC/MD
Sample : L2230-54
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1-9-17

Quant Time: Apr 16 07:48:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

