

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041520\
 Data File : VN060972.D
 Acq On : 15 Apr 2020 19:16
 Operator : JC/MD
 Sample : L2255-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-27-26-COMP

Quant Time: Apr 16 07:50:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	136617	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	231340	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	220671	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	98538	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	90959	48.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.00%	
35) Dibromofluoromethane	7.56	113	69179	49.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.90%	
50) Toluene-d8	10.08	98	288377	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
62) 4-Bromofluorobenzene	12.40	95	108978	52.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.08%	
Target Compounds						
16) Acetone	3.78	43	13135	19.772	ug/l	97
20) Methylene Chloride	4.52	84	2524	1.544	ug/l	96
43) Isopropyl Acetate	8.14	43	49321	12.356	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041520\
Data File : VN060972.D
Acq On : 15 Apr 2020 19:16
Operator : JC/MD
Sample : L2255-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-27-26-COMP

Quant Time: Apr 16 07:50:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

