

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041520\
 Data File : VN060979.D
 Acq On : 15 Apr 2020 22:17
 Operator : JC/MD
 Sample : VN0415WBL02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0415WBL02

Quant Time: Apr 16 08:06:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	136965	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	231801	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	220311	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	98076	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	93086	49.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.02%	
35) Dibromofluoromethane	7.56	113	69655	49.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.40%	
50) Toluene-d8	10.08	98	289431	50.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.10%	
62) 4-Bromofluorobenzene	12.39	95	108408	51.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.34%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041520\
Data File : VN060979.D
Acq On : 15 Apr 2020 22:17
Operator : JC/MD
Sample : VN0415WBL02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0415WBL02

Quant Time: Apr 16 08:06:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

