

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041520\
 Data File : VN060982.D
 Acq On : 15 Apr 2020 23:25
 Operator : JC/MD
 Sample : L2123-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-04-142004

Quant Time: Apr 16 08:16:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	127640	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	215438	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	199445	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	87507	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	84982	48.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.00%	
35) Dibromofluoromethane	7.56	113	63441	48.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.40%	
50) Toluene-d8	10.08	98	267595	50.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.58%	
62) 4-Bromofluorobenzene	12.40	95	97487	49.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.98%	
Target Compounds						
16) Acetone	3.78	43	15844	25.527	ug/l	96
20) Methylene Chloride	4.52	84	4929	3.228	ug/l	99
25) 2-Butanone	6.81	43	5804	5.779	ug/l	95
43) Isopropyl Acetate	8.14	43	52986	14.253	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041520\
Data File : VN060982.D
Acq On : 15 Apr 2020 23:25
Operator : JC/MD
Sample : L2123-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-04-142004

Quant Time: Apr 16 08:16:28 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

