

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041520\
 Data File : VN060983.D
 Acq On : 15 Apr 2020 23:48
 Operator : JC/MD
 Sample : L2252-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CHASSIS-WASH-SEDIMENT-PIT

Quant Time: Apr 16 08:19:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	136101	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	230781	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	220972	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	104378	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.00	65	92456	49.49	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	98.98%	
35) Dibromofluoromethane	7.56	113	69323	49.68	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	99.36%	
50) Toluene-d8	10.08	98	293537	51.50	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	103.00%	
62) 4-Bromofluorobenzene	12.39	95	113318	54.25	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	108.50%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.78	43	40534	61.247	ug/l	99
18) Methyl Acetate	4.30	43	3281	1.761	ug/l #	83
20) Methylene Chloride	4.52	84	4104	2.521	ug/l	85
25) 2-Butanone	6.79	43	23620	22.056	ug/l	94
30) Chloroform	7.34	83	9507	3.322	ug/l	98
43) Isopropyl Acetate	8.14	43	37230	9.349	ug/l #	89
51) 4-Methyl-2-Pentanone	9.97	43	113872	50.252	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041520\
Data File : VN060983.D
Acq On : 15 Apr 2020 23:48
Operator : JC/MD
Sample : L2252-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CHASSIS-WASH-SEDIMENT-PIT

Quant Time: Apr 16 08:19:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

