

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041520\
 Data File : VN060994.D
 Acq On : 16 Apr 2020 3:58
 Operator : JC/MD
 Sample : L2249-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-1

Quant Time: Apr 16 08:38:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	144267	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	245230	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	229119	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	101684	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	96577	48.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.54%	
35) Dibromofluoromethane	7.56	113	73474	49.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.10%	
50) Toluene-d8	10.08	98	304075	50.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.40%	
62) 4-Bromofluorobenzene	12.39	95	112784	50.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.62%	
Target Compounds						
16) Acetone	3.78	43	2358	3.361	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041520\
Data File : VN060994.D
Acq On : 16 Apr 2020 3:58
Operator : JC/MD
Sample : L2249-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-1

Quant Time: Apr 16 08:38:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

