

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041520\
 Data File : VN060999.D
 Acq On : 16 Apr 2020 5:51
 Operator : JC/MD
 Sample : L2249-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-6

Quant Time: Apr 16 08:21:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	141936	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	241935	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	226122	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	101996	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	94365	48.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.86%	
35) Dibromofluoromethane	7.56	113	71549	48.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.82%	
50) Toluene-d8	10.08	98	297843	49.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.68%	
62) 4-Bromofluorobenzene	12.39	95	112581	51.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.82%	
Target Compounds						
16) Acetone	3.79	43	2253	3.264	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041520\
Data File : VN060999.D
Acq On : 16 Apr 2020 5:51
Operator : JC/MD
Sample : L2249-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-6

Quant Time: Apr 16 08:21:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

