

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041521\
 Data File : VN066613.D
 Acq On : 15 Apr 2021 15:51
 Operator : JC/MD
 Sample : VN0415WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0415WBL01

Quant Time: Apr 16 07:51:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 15 19:16:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.586	168	238978	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.514	114	382569	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.349	117	348341	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	137736	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.950	65	178345	49.47	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.94%
35) Dibromofluoromethane	7.508	113	137853	51.96	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.92%
50) Toluene-d8	10.026	98	546268	52.53	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.06%
62) 4-Bromofluorobenzene	12.346	95	174559	47.74	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.48%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041521\
Data File : VN066613.D
Acq On : 15 Apr 2021 15:51
Operator : JC/MD
Sample : VN0415WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0415WBL01

Quant Time: Apr 16 07:51:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 15 19:16:35 2021
Response via : Initial Calibration

