

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041521\
 Data File : VN066617.D
 Acq On : 15 Apr 2021 17:31
 Operator : JC/MD
 Sample : PB135761TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135761TB

Quant Time: Apr 16 07:53:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 15 19:16:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.588	168	226388	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	360869	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	349607	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	138005	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	171374	50.18	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.36%	
35) Dibromofluoromethane	7.510	113	130963	52.33	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 104.66%	
50) Toluene-d8	10.029	98	528706	53.90	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 107.80%	
62) 4-Bromofluorobenzene	12.346	95	173723	50.37	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 100.74%	
Target Compounds						
						Qvalue
16) Acetone	3.753	43	19828	8.01	ug/l	97
20) Methylene Chloride	4.466	84	14232	4.41	ug/l	92
25) 2-Butanone	6.754	43	10266	5.10	ug/l	94
43) Isopropyl Acetate	8.095	43	26267	3.81	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041521\
Data File : VN066617.D
Acq On : 15 Apr 2021 17:31
Operator : JC/MD
Sample : PB135761TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135761TB

Quant Time: Apr 16 07:53:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 15 19:16:35 2021
Response via : Initial Calibration

