

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041521\
 Data File : VN066619.D
 Acq On : 15 Apr 2021 18:18
 Operator : JC/MD
 Sample : M2033-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FM5210

Manual Integrations
 APPROVED

MMDadoda
 4/16/2021 1:10:51 PM

Quant Time: Apr 16 08:56:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 15 19:16:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.588	168	231981	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.516	114	369498	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	345910	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	142896	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.950	65	177314	50.67	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 101.34%			
35) Dibromofluoromethane	7.510	113	136352	53.21	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 106.42%			
50) Toluene-d8	10.029	98	534267	53.20	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 106.40%			
62) 4-Bromofluorobenzene	12.349	95	173876	49.24	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 98.48%			
Target Compounds						Qvalue
16) Acetone	3.758	43	15951	4.40	ug/l	91
18) Methyl Acetate	4.257	43	5910	1.57	ug/l #	83
52) Toluene	10.093	92	271709	38.66	ug/l	99
80) 1,3,5-Trimethylbenzene	12.689	105	4329m	0.47	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041521\
Data File : VN066619.D
Acq On : 15 Apr 2021 18:18
Operator : JC/MD
Sample : M2033-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FM5210

Manual Integrations
APPROVED

MMDadoda
4/16/2021 1:10:51 PM

Quant Time: Apr 16 08:56:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 15 19:16:35 2021
Response via : Initial Calibration

