

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041521\
 Data File : VN066623.D
 Acq On : 15 Apr 2021 19:53
 Operator : JC/MD
 Sample : M2030-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ETGI-304

Quant Time: Apr 16 07:54:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 15 19:16:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.588	168	239775	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.516	114	376469	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	377985	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	158545	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.950	65	178674	49.40	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.80%
35) Dibromofluoromethane	7.510	113	137806	52.78	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	105.56%
50) Toluene-d8	10.029	98	564806	55.20	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	110.40%
62) 4-Bromofluorobenzene	12.346	95	192541	53.51	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.02%
Target Compounds						
					Qvalue	
16) Acetone	3.756	43	16749	4.61	ug/l	93
18) Methyl Acetate	4.257	43	5137	1.32	ug/l #	66
20) Methylene Chloride	4.466	84	14211	4.16	ug/l	88
43) Isopropyl Acetate	8.095	43	23694	3.30	ug/l #	86
95) Naphthalene	15.076	128	50754	6.27	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041521\
Data File : VN066623.D
Acq On : 15 Apr 2021 19:53
Operator : JC/MD
Sample : M2030-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ETGI-304

Quant Time: Apr 16 07:54:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 15 19:16:35 2021
Response via : Initial Calibration

