

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041522\
 Data File : VN072009.D
 Acq On : 15 Apr 2022 18:26
 Operator : JC\MD
 Sample : N2441-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT-TB-20220413

Quant Time: Apr 15 23:29:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032422W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Mar 24 12:03:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.657	128	81805	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.963	114	430082	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.739	117	365333	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.434	65	162444	28.184	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	93.933%
60) 4-Bromofluorobenzene	12.727	95	122342	22.585	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	75.267%
63) Toluene-d8	10.439	98	503196	30.342	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.133%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041522\
Data File : VN072009.D
Acq On : 15 Apr 2022 18:26
Operator : JC\MD
Sample : N2441-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT-TB-20220413

Quant Time: Apr 15 23:29:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032422W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Mar 24 12:03:44 2022
Response via : Initial Calibration

