

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041522\
 Data File : VN072014.D
 Acq On : 15 Apr 2022 20:27
 Operator : JC\MD
 Sample : PB144101ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB144101ZHE#02

Quant Time: Apr 16 00:08:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	627031	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1068303	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1223597	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	426601	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	337987	47.216	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.440%
35) Dibromofluoromethane	8.016	113	301341	47.936	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.880%
50) Toluene-d8	10.439	98	1427417	55.507	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	111.020%
62) 4-Bromofluorobenzene	12.727	95	485812	58.412	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	116.820%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	18267	5.971	ug/l #	88
20) Methylene Chloride	5.093	84	11314	1.755	ug/l	94
43) Isopropyl Acetate	8.557	43	81194	5.080	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041522\
Data File : VN072014.D
Acq On : 15 Apr 2022 20:27
Operator : JC\MD
Sample : PB144101ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB144101ZHE#02

Quant Time: Apr 16 00:08:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

