

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041522\
 Data File : VN072019.D
 Acq On : 15 Apr 2022 22:27
 Operator : JC\MD
 Sample : PB144101ZHE#07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB144101ZHE#07

Quant Time: Apr 16 00:09:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	731127	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1253956	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1397757	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	495273	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	390162	46.745	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	93.480%		
35) Dibromofluoromethane	8.016	113	349147	47.317	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.640%		
50) Toluene-d8	10.439	98	1647392	54.576	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	109.160%		
62) 4-Bromofluorobenzene	12.727	95	567313	58.113	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	116.220%		
Target Compounds						
20) Methylene Chloride	5.081	84	11168	1.485	ug/l #	83
43) Isopropyl Acetate	8.557	43	113464	6.048	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041522\
Data File : VN072019.D
Acq On : 15 Apr 2022 22:27
Operator : JC\MD
Sample : PB144101ZHE#07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB144101ZHE#07

Quant Time: Apr 16 00:09:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

