

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041522\
 Data File : VN072020.D
 Acq On : 15 Apr 2022 22:51
 Operator : JC\MD
 Sample : PB144101ZHE#08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB144101ZHE#08

Quant Time: Apr 16 00:09:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	679569	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1171913	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1292267	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	448299	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	367700	47.396	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.800%
35) Dibromofluoromethane	8.016	113	328600	47.651	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.300%
50) Toluene-d8	10.439	98	1527985	54.164	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.320%
62) 4-Bromofluorobenzene	12.727	95	520121	57.008	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	114.020%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	16845	5.081	ug/l #	72
20) Methylene Chloride	5.093	84	10407	1.489	ug/l #	91
43) Isopropyl Acetate	8.557	43	107160	6.112	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041522\
Data File : VN072020.D
Acq On : 15 Apr 2022 22:51
Operator : JC\MD
Sample : PB144101ZHE#08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB144101ZHE#08

Quant Time: Apr 16 00:09:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

