

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041522\
 Data File : VN072021.D
 Acq On : 15 Apr 2022 23:15
 Operator : JC\MD
 Sample : PB144101ZHE#09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB144101ZHE#09

Quant Time: Apr 16 00:09:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	710448	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1215080	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1304298	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	480069	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	384701	47.432	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.860%		
35) Dibromofluoromethane	8.016	113	343347	48.020	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	96.040%		
50) Toluene-d8	10.439	98	1576637	53.904	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	107.800%		
62) 4-Bromofluorobenzene	12.727	95	527671	55.781	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	111.560%		
Target Compounds						
20) Methylene Chloride	5.099	84	10692	1.463	ug/l	91
43) Isopropyl Acetate	8.563	43	110361	6.071	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041522\
Data File : VN072021.D
Acq On : 15 Apr 2022 23:15
Operator : JC\MD
Sample : PB144101ZHE#09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB144101ZHE#09

Quant Time: Apr 16 00:09:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

