

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041522\
 Data File : VN072024.D
 Acq On : 16 Apr 2022 00:28
 Operator : JC\MD
 Sample : PB144101ZHE#12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB144101ZHE#12

Quant Time: Apr 16 01:12:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	642986	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1110897	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1207519	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	454510	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	347825	47.385	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.760%
35) Dibromofluoromethane	8.016	113	312898	47.866	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.740%
50) Toluene-d8	10.439	98	1430833	53.506	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.020%
62) 4-Bromofluorobenzene	12.727	95	503634	58.233	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	116.460%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	16987	5.415	ug/l	98
20) Methylene Chloride	5.093	84	10999	1.663	ug/l	92
43) Isopropyl Acetate	8.563	43	158425	9.533	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041522\
Data File : VN072024.D
Acq On : 16 Apr 2022 00:28
Operator : JC\MD
Sample : PB144101ZHE#12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB144101ZHE#12

Quant Time: Apr 16 01:12:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

