


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041522\
 Data File : VN072001.D
 Acq On : 15 Apr 2022 14:12
 Operator : JC\MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Title : SW846 8260
 Last Update : Thu Mar 31 01:57:54 2022



AutoFind: Scans 1825, 1826, 1827; Background Corrected with Scan 1817

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.8	8905	PASS
75	95	30	60	48.0	26971	PASS
95	95	100	100	100.0	56211	PASS
96	95	5	9	7.1	3964	PASS
173	174	0.00	2	1.1	580	PASS
174	95	50	100	91.0	51141	PASS
175	174	5	9	7.2	3673	PASS
176	174	95	101	96.3	49237	PASS
177	176	5	9	6.2	3038	PASS