

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041524\
 Data File : VN081740.D
 Acq On : 15 Apr 2024 16:12
 Operator : JC\MD
 Sample : VN0415MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0415MBL01

Quant Time: Apr 15 23:28:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041524W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 15:37:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	205148	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	370871	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	332809	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	132682	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	150965	52.296	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.600%
35) Dibromofluoromethane	8.165	113	124086	50.667	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.340%
50) Toluene-d8	10.571	98	453852	50.304	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.600%
62) 4-Bromofluorobenzene	12.853	95	141199	45.798	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	91.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041524\
Data File : VN081740.D
Acq On : 15 Apr 2024 16:12
Operator : JC\MD
Sample : VN0415MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0415MBL01

Quant Time: Apr 15 23:28:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041524W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 15:37:39 2024
Response via : Initial Calibration

