

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041525\
 Data File : VN086281.D
 Acq On : 15 Apr 2025 13:51
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 04/16/2025
 Supervised By : Semsettin Yesilyurt 04/16/2025

Quant Time: Apr 16 03:57:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 03:42:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	223468	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	413969	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	367304	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	173703	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	160758	49.609	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.220%		
35) Dibromofluoromethane	8.165	113	88836	46.238	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.480%		
50) Toluene-d8	10.565	98	517785	50.426	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	100.860%		
62) 4-Bromofluorobenzene	12.847	95	190049	50.744	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	101.480%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	132017	49.836	ug/l	100
3) Chloromethane	2.359	50	168565	43.783	ug/l	100
4) Vinyl Chloride	2.518	62	176686	48.092	ug/l	100
5) Bromomethane	2.953	94	79546	48.135	ug/l	100
6) Chloroethane	3.118	64	111725	45.437	ug/l	100
7) Trichlorofluoromethane	3.501	101	192876	47.395	ug/l	100
8) Diethyl Ether	3.959	74	83664	47.092	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.371	101	117478	47.719	ug/l	100
10) Methyl Iodide	4.589	142	133779	49.427	ug/l	100
11) Tert butyl alcohol	5.512	59	138743	234.659	ug/l	100
12) 1,1-Dichloroethene	4.342	96	123106	46.805	ug/l	100
13) Acrolein	4.177	56	68075	246.717	ug/l	100
14) Allyl chloride	5.024	41	207525	44.409	ug/l	100
15) Acrylonitrile	5.712	53	354632	242.632	ug/l	100
16) Acetone	4.424	43	280935	250.761	ug/l	100
17) Carbon Disulfide	4.712	76	348133	44.205	ug/l	100
18) Methyl Acetate	5.018	43	174600	44.886	ug/l	100
19) Methyl tert-butyl Ether	5.789	73	449127	46.453	ug/l	100
20) Methylene Chloride	5.271	84	138293	46.163	ug/l	100
21) trans-1,2-Dichloroethene	5.789	96	128404	46.696	ug/l	100
22) Diisopropyl ether	6.665	45	471385	46.345	ug/l	100
23) Vinyl Acetate	6.600	43	1657348	233.153	ug/l	100
24) 1,1-Dichloroethane	6.565	63	253896	47.310	ug/l	100
25) 2-Butanone	7.477	43	475281	235.633	ug/l	100
26) 2,2-Dichloropropane	7.483	77	219735	45.729	ug/l	100
27) cis-1,2-Dichloroethene	7.483	96	159460	46.722	ug/l	100
28) Bromochloromethane	7.806	49	110726	48.660	ug/l	100
29) Tetrahydrofuran	7.836	42	313920	232.739	ug/l	100
30) Chloroform	7.965	83	246843	46.823	ug/l	100
31) Cyclohexane	8.253	56	238117	45.343	ug/l	100
32) 1,1,1-Trichloroethane	8.165	97	212021	47.013	ug/l	100
36) 1,1-Dichloropropene	8.365	75	183246	47.752	ug/l	100
37) Ethyl Acetate	7.553	43	184834	45.750	ug/l	100
38) Carbon Tetrachloride	8.359	117	175094	47.912	ug/l	100
39) Methylcyclohexane	9.594	83	216725	47.515	ug/l	100
40) Benzene	8.600	78	583316	47.443	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041525\
 Data File : VN086281.D
 Acq On : 15 Apr 2025 13:51
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By :John Carlone 04/16/2025
 Supervised By :Semsettin Yesilyurt 04/16/2025

Quant Time: Apr 16 03:57:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 03:42:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.771	41	109214	46.643	ug/l	100
42) 1,2-Dichloroethane	8.665	62	183854	46.892	ug/l	100
43) Isopropyl Acetate	8.683	43	358327	49.550	ug/l	100
44) Trichloroethene	9.347	130	139993	48.039	ug/l	100
45) 1,2-Dichloropropane	9.618	63	144496	48.011	ug/l	100
46) Dibromomethane	9.706	93	93965	49.002	ug/l	100
47) Bromodichloromethane	9.882	83	196793	47.516	ug/l	100
48) Methyl methacrylate	9.677	41	163719	46.146	ug/l	100
49) 1,4-Dioxane	9.694	88	53851	892.337	ug/l	100
51) 4-Methyl-2-Pentanone	10.441	43	1000994	241.959	ug/l	100
52) Toluene	10.629	92	369349	48.144	ug/l	100
53) t-1,3-Dichloropropene	10.829	75	226518	48.996	ug/l	100
54) cis-1,3-Dichloropropene	10.306	75	238764	47.021	ug/l	100
55) 1,1,2-Trichloroethane	11.012	97	131721	47.721	ug/l	100
56) Ethyl methacrylate	10.871	69	239654	47.235	ug/l	100
57) 1,3-Dichloropropane	11.159	76	233548	47.689	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	594475	246.859	ug/l	100
59) 2-Hexanone	11.194	43	734660	239.414	ug/l	100
60) Dibromochloromethane	11.359	129	146318	48.274	ug/l	100
61) 1,2-Dibromoethane	11.465	107	134281	48.213	ug/l	100
64) Tetrachloroethene	11.100	164	136116	48.177	ug/l	100
65) Chlorobenzene	11.888	112	385849	47.072	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.959	131	127627	47.197	ug/l	100
67) Ethyl Benzene	11.959	91	704474	47.627	ug/l	100
68) m/p-Xylenes	12.065	106	533742	96.302	ug/l	100
69) o-Xylene	12.394	106	261904	47.783	ug/l	100
70) Styrene	12.406	104	447486	48.998	ug/l	100
71) Bromoform	12.576	173	95822	47.065	ug/l #	100
73) Isopropylbenzene	12.694	105	647490	45.568	ug/l	100
74) N-amyl acetate	12.488	43	308916	43.670	ug/l	100
75) 1,1,2,2-Tetrachloroethane	12.935	83	179865	44.025	ug/l	100
76) 1,2,3-Trichloropropane	12.994	75	181010m	44.357	ug/l	
77) Bromobenzene	12.976	156	147720	46.919	ug/l	100
78) n-propylbenzene	13.035	91	779530	46.556	ug/l	100
79) 2-Chlorotoluene	13.123	91	479235	45.740	ug/l	100
80) 1,3,5-Trimethylbenzene	13.170	105	535465	46.042	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	80078m	46.919	ug/l	
82) 4-Chlorotoluene	13.218	91	488049	46.893	ug/l	100
83) tert-Butylbenzene	13.435	119	452999	44.954	ug/l	100
84) 1,2,4-Trimethylbenzene	13.476	105	550727	46.521	ug/l	100
85) sec-Butylbenzene	13.612	105	653679	46.767	ug/l	100
86) p-Isopropyltoluene	13.723	119	548528	47.611	ug/l	100
87) 1,3-Dichlorobenzene	13.729	146	279645	47.249	ug/l	100
88) 1,4-Dichlorobenzene	13.812	146	280378	47.032	ug/l	100
89) n-Butylbenzene	14.053	91	492941	48.271	ug/l	100
90) Hexachloroethane	14.329	117	90009	46.447	ug/l	100
91) 1,2-Dichlorobenzene	14.106	146	270731	47.181	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	38625	46.887	ug/l	100
93) 1,2,4-Trichlorobenzene	15.388	180	132919	48.375	ug/l	100
94) Hexachlorobutadiene	15.500	225	48694	46.947	ug/l	100
95) Naphthalene	15.635	128	464685	47.708	ug/l	100
96) 1,2,3-Trichlorobenzene	15.835	180	122910	47.249	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041525\
Data File : VN086281.D
Acq On : 15 Apr 2025 13:51
Operator : JC\MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/16/2025
Supervised By :Semsettin Yesilyurt 04/16/2025

Quant Time: Apr 16 03:57:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 03:42:50 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

