

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041619\
 Data File : VN055071.D
 Acq On : 16 Apr 2019 12:03
 Operator : JC/SP
 Sample : VN0416WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0416WBL01

Quant Time: Apr 16 12:23:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 12:28:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	553014	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	853567	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	701522	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	217279	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.02	65	332899	45.29	ug/l	0.00
Spiked Amount						
			Recovery	=	90.58%	
35) Dibromofluoromethane	7.59	113	269578	47.26	ug/l	0.00
Spiked Amount						
			Recovery	=	94.52%	
50) Toluene-d8	10.09	98	1011410	49.54	ug/l	0.00
Spiked Amount						
			Recovery	=	99.08%	
62) 4-Bromofluorobenzene	12.40	95	289696	43.65	ug/l	0.00
Spiked Amount						
			Recovery	=	87.30%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041619\
Data File : VN055071.D
Acq On : 16 Apr 2019 12:03
Operator : JC/SP
Sample : VN0416WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0416WBL01

Quant Time: Apr 16 12:23:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 12:28:28 2019
Response via : Initial Calibration

