

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041619\
 Data File : VN055081.D
 Acq On : 16 Apr 2019 16:26
 Operator : JC/SP
 Sample : K2422-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP1-D

Quant Time: Apr 17 02:28:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 12:28:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	521086	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	833730	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	703855	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	223731	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	323616	46.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.44%	
35) Dibromofluoromethane	7.59	113	261176	46.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.74%	
50) Toluene-d8	10.09	98	983450	49.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.64%	
62) 4-Bromofluorobenzene	12.40	95	291375	44.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.88%	
Target Compounds						
20) Methylene Chloride	4.54	84	13141	1.988	ug/l #	88
43) Isopropyl Acetate	8.17	43	13111	1.330	ug/l	98
95) Naphthalene	15.13	128	3188	3.453	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041619\
Data File : VN055081.D
Acq On : 16 Apr 2019 16:26
Operator : JC/SP
Sample : K2422-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP1-D

Quant Time: Apr 17 02:28:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 12:28:28 2019
Response via : Initial Calibration

