

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041619\
 Data File : VN055082.D
 Acq On : 16 Apr 2019 16:53
 Operator : JC/SP
 Sample : K2422-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP1-S

Quant Time: Apr 17 02:31:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 12:28:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	507833	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	825107	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	687233	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	225703	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	328591	48.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.36%	
35) Dibromofluoromethane	7.59	113	263000	47.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.38%	
50) Toluene-d8	10.09	98	975867	49.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.90%	
62) 4-Bromofluorobenzene	12.40	95	285091	44.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.86%	
Target Compounds						
20) Methylene Chloride	4.54	84	89698	13.927	ug/l	95
43) Isopropyl Acetate	8.17	43	12789	1.311	ug/l	98
84) 1,2,4-Trimethylbenzene	13.04	105	29641	1.935	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041619\
Data File : VN055082.D
Acq On : 16 Apr 2019 16:53
Operator : JC/SP
Sample : K2422-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP1-S

Quant Time: Apr 17 02:31:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 12:28:28 2019
Response via : Initial Calibration

