

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041619\
 Data File : VN055071.D
 Acq On : 16 Apr 2019 12:03
 Operator : JC/SP
 Sample : VN0416WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0416WBL01

Quant Time: Apr 16 12:23:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 15 12:28:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	553014	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	853567	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	701522	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	217279	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.02	65	332899	45.29	ug/l	0.00
Spiked Amount						
			Recovery	=	90.58%	
35) Dibromofluoromethane	7.59	113	269578	47.26	ug/l	0.00
Spiked Amount						
			Recovery	=	94.52%	
50) Toluene-d8	10.09	98	1011410	49.54	ug/l	0.00
Spiked Amount						
			Recovery	=	99.08%	
62) 4-Bromofluorobenzene	12.40	95	289696	43.65	ug/l	0.00
Spiked Amount						
			Recovery	=	87.30%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041619\
Data File : VN055071.D
Acq On : 16 Apr 2019 12:03
Operator : JC/SP
Sample : VN0416WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0416WBL01

Quant Time: Apr 16 12:23:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041519W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 15 12:28:28 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.66 min Scan# 1832

Delta R.T. -0.00 min

Lab File: VN055071.D

Acq: 16 Apr 2019 12:03

Instrument :

MSVOA_N

ClientSampled :


VN0416WBL01

Tot Ion:168 Resp: 553014

Ion Ratio Lower Upper




168 100

99 52.8 41.9 62.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN



#33

1,2-Dichloroethane-d4

Concen: 45.287 ug/l

RT: 8.02 min Scan# 1944

Delta R.T. -0.00 min

Lab File: VN055071.D


Acq: 16 Apr 2019 12:03

Tgt Ion: 65 Resp: 332899

Ion Ratio Lower Upper



65 100


67 52.4 0.0 106.4



Abundance Ion 64.90 (64.60 to 65.60): VN

Ion 66.90 (66.60 to 67.60): VN





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.58 min Scan# 2118

Delta R.T. 0.00 min

Lab File: VN055071.D

Acq: 16 Apr 2019 12:03

Instrument :

MSVOA_N

ClientSampleId :

VN0416WBL01


Tot Ion:114 Resp: 853567

Ion Ratio Lower Upper

114 100

63 22.0 0.0 45.2




88 15.5 0.0 31.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN0

Ion 87.90 (87.60 to 88.60): VN0



#35

Dibromofluoromethane

Concen: 47.257 ug/l

RT: 7.59 min Scan# 1809

Delta R.T. -0.00 min

Lab File: VN055071.D

Acq: 16 Apr 2019 12:03


Tgt Ion:113 Resp: 269578

Ion Ratio Lower Upper

113 100

111 103.2 82.1 123.1



192 21.2 16.5 24.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 49.538 ug/l

RT: 10.09 min Scan# 2586

Delta R.T. 0.00 min

Lab File: VN055071.D

Acq: 16 Apr 2019 12:03

Instrument :

MSVOA_N

ClientSampleId :





VN0416WBL01

Tot Ion: 98 Resp: 1011410

Ion Ratio Lower Upper

98 100

100 64.0 51.1 76.7



#62

4-Bromofluorobenzene

Concen: 43.646 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. 0.00 min

Lab File: VN055071.D

Acq: 16 Apr 2019 12:03




Tgt Ion: 95 Resp: 289696


Ion Ratio Lower Upper

95 100

174 86.1 0.0 174.4

176 82.6 0.0 166.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. -0.00 min
Lab File: VN055071.D
Acq: 16 Apr 2019 12:03

Instrument :
MSVOA_N
ClientSampled :
VN0416WBL01

Tot Ion:117 Resp: 701522
Ion Ratio Lower Upper
117 100
82 57.2 44.6 66.8
119 32.6 25.5 38.3




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.34 min Scan# 3598
Delta R.T. 0.00 min
Lab File: VN055071.D
Acq: 16 Apr 2019 12:03

Tgt Ion:152 Resp: 217279
Ion Ratio Lower Upper
152 100
115 57.9 28.5 85.6
150 160.1 0.0 347.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

