

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041620\
 Data File : VN061012.D
 Acq On : 16 Apr 2020 15:22
 Operator : JC/MD
 Sample : VN0416WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0416WBL01

Quant Time: Apr 17 04:17:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 17 03:49:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	153677	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	260226	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	235686	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	94065	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.00	65	101545	51.65	ug/l	0.00
Spiked Amount						
			Recovery	=	103.30%	
35) Dibromofluoromethane	7.56	113	77829	51.21	ug/l	0.00
Spiked Amount						
			Recovery	=	102.42%	
50) Toluene-d8	10.08	98	318138	55.40	ug/l	0.00
Spiked Amount						
			Recovery	=	110.80%	
62) 4-Bromofluorobenzene	12.39	95	111655	49.61	ug/l	0.00
Spiked Amount						
			Recovery	=	99.22%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041620\
Data File : VN061012.D
Acq On : 16 Apr 2020 15:22
Operator : JC/MD
Sample : VN0416WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0416WBL01

Quant Time: Apr 17 04:17:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 17 03:49:39 2020
Response via : Initial Calibration

