

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041620\
 Data File : VN061015.D
 Acq On : 16 Apr 2020 16:41
 Operator : JC/MD
 Sample : L2259-01DL 20X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CHASSIS-BUS-WASHDL

Quant Time: Apr 17 04:31:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 17 03:49:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	158712	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	266248	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	242395	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	101643	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	103376	50.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.82%	
35) Dibromofluoromethane	7.56	113	78584	50.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.08%	
50) Toluene-d8	10.08	98	324397	55.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.42%	
62) 4-Bromofluorobenzene	12.39	95	117094	50.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.70%	
Target Compounds						
20) Methylene Chloride	4.53	84	3088	1.582	ug/l	93
30) Chloroform	7.35	83	313854	91.284	ug/l	99
95) Naphthalene	15.13	128	6757	1.210	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041620\
Data File : VN061015.D
Acq On : 16 Apr 2020 16:41
Operator : JC/MD
Sample : L2259-01DL 20X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CHASSIS-BUS-WASHDL

Quant Time: Apr 17 04:31:06 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 17 03:49:39 2020
Response via : Initial Calibration

