

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041620\
 Data File : VN061019.D
 Acq On : 16 Apr 2020 18:12
 Operator : JC/MD
 Sample : PB128250TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB128250TB

Quant Time: Apr 17 04:48:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 17 03:49:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	155127	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	262858	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	252242	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	113821	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	104188	52.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.98%	
35) Dibromofluoromethane	7.56	113	79121	51.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.08%	
50) Toluene-d8	10.08	98	328756	56.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.36%	
62) 4-Bromofluorobenzene	12.39	95	124720	54.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.72%	
Target Compounds						
16) Acetone	3.78	43	21896	25.064	ug/l	97
43) Isopropyl Acetate	8.14	43	55871	11.664	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041620\
Data File : VN061019.D
Acq On : 16 Apr 2020 18:12
Operator : JC/MD
Sample : PB128250TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB128250TB

Quant Time: Apr 17 04:48:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 17 03:49:39 2020
Response via : Initial Calibration

