

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041620\
 Data File : VN061020.D
 Acq On : 16 Apr 2020 18:35
 Operator : JC/MD
 Sample : L2284-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-11-SA

Quant Time: Apr 17 04:53:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 17 03:49:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	177666	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	305792	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	286190	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	126568	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	119442	52.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.10%	
35) Dibromofluoromethane	7.56	113	90153	50.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.96%	
50) Toluene-d8	10.08	98	376085	55.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.46%	
62) 4-Bromofluorobenzene	12.39	95	140112	52.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.96%	
Target Compounds						
16) Acetone	3.78	43	100582	100.530	ug/l	97
20) Methylene Chloride	4.52	84	7992	3.659	ug/l	99
25) 2-Butanone	6.80	43	12864	8.391	ug/l	99
43) Isopropyl Acetate	8.14	43	47484	8.521	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041620\
Data File : VN061020.D
Acq On : 16 Apr 2020 18:35
Operator : JC/MD
Sample : L2284-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-11-SA

Quant Time: Apr 17 04:53:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 17 03:49:39 2020
Response via : Initial Calibration

