

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041620\
 Data File : VN061021.D
 Acq On : 16 Apr 2020 18:57
 Operator : JC/MD
 Sample : L2284-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-11-DA

Quant Time: Apr 17 04:57:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 17 03:49:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	163343	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	278572	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	264925	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	119635	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	111102	53.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.32%	
35) Dibromofluoromethane	7.56	113	83020	51.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.06%	
50) Toluene-d8	10.08	98	347239	56.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.98%	
62) 4-Bromofluorobenzene	12.39	95	132627	55.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.10%	
Target Compounds						
8) Diethyl Ether	3.38	74	1516	1.273	ug/l	93
16) Acetone	3.78	43	99698	108.384	ug/l	97
20) Methylene Chloride	4.52	84	8600	4.282	ug/l	99
25) 2-Butanone	6.80	43	13383	9.494	ug/l	96
43) Isopropyl Acetate	8.14	43	44194	8.706	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041620\
Data File : VN061021.D
Acq On : 16 Apr 2020 18:57
Operator : JC/MD
Sample : L2284-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-11-DA

Quant Time: Apr 17 04:57:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 17 03:49:39 2020
Response via : Initial Calibration

