

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041620\
 Data File : VN061026.D
 Acq On : 16 Apr 2020 20:51
 Operator : JC/MD
 Sample : L2298-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SOIL-CUTTINGS

Quant Time: Apr 17 06:14:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 17 03:49:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	150481	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	257726	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	242273	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	105790	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	101633	52.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.58%	
35) Dibromofluoromethane	7.56	113	76389	50.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.50%	
50) Toluene-d8	10.08	98	323126	56.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.62%	
62) 4-Bromofluorobenzene	12.39	95	118242	53.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.10%	
Target Compounds						
16) Acetone	3.78	43	75849	89.505	ug/l	99
20) Methylene Chloride	4.52	84	5197	2.809	ug/l	90
25) 2-Butanone	6.80	43	9016	6.943	ug/l	96
43) Isopropyl Acetate	8.14	43	41101	8.751	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041620\
Data File : VN061026.D
Acq On : 16 Apr 2020 20:51
Operator : JC/MD
Sample : L2298-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SOIL-CUTTINGS

Quant Time: Apr 17 06:14:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 17 03:49:39 2020
Response via : Initial Calibration

