

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN041620\
 Data File : VN061030.D
 Acq On : 16 Apr 2020 22:21
 Operator : JC/MD
 Sample : L2304-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-B2-VOA

Quant Time: Apr 17 06:23:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 17 03:49:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	159640	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	273224	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	260609	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	118644	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	109355	53.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.08%	
35) Dibromofluoromethane	7.56	113	81801	51.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.54%	
50) Toluene-d8	10.08	98	340497	56.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.94%	
62) 4-Bromofluorobenzene	12.39	95	129326	54.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.46%	
Target Compounds						
16) Acetone	3.78	43	26813	29.825	ug/l	99
20) Methylene Chloride	4.52	84	2406	1.226	ug/l	93
43) Isopropyl Acetate	8.14	43	56723	11.392	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN041620\
Data File : VN061030.D
Acq On : 16 Apr 2020 22:21
Operator : JC/MD
Sample : L2304-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-B2-VOA

Quant Time: Apr 17 06:23:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N041620W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 17 03:49:39 2020
Response via : Initial Calibration

