

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041621\
 Data File : VN066627.D
 Acq On : 16 Apr 2021 15:34
 Operator : JC/MD
 Sample : VN0416WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0416WBL01

Quant Time: Apr 17 05:16:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 15 19:16:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.588	168	245660	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.516	114	389645	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	359254	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	143442	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.950	65	179723	48.50	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.00%
35) Dibromofluoromethane	7.508	113	139860	51.76	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.52%
50) Toluene-d8	10.026	98	551489	52.07	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.14%
62) 4-Bromofluorobenzene	12.346	95	171621	46.09	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.18%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041621\
Data File : VN066627.D
Acq On : 16 Apr 2021 15:34
Operator : JC/MD
Sample : VN0416WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0416WBL01

Quant Time: Apr 17 05:16:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 15 19:16:35 2021
Response via : Initial Calibration

