

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041621\
 Data File : VN066630.D
 Acq On : 16 Apr 2021 16:56
 Operator : JC/MD
 Sample : PB135768TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135768TB

Quant Time: Apr 17 05:17:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 15 19:16:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	259553	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.516	114	411448	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	399524	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	155668	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.950	65	190321	48.61	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.22%
35) Dibromofluoromethane	7.511	113	148813	52.15	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.30%
50) Toluene-d8	10.029	98	600785	53.72	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.44%
62) 4-Bromofluorobenzene	12.346	95	198906	50.58	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.16%
Target Compounds						
					Qvalue	
16) Acetone	3.756	43	26258	10.62	ug/l	95
20) Methylene Chloride	4.467	84	12886	3.48	ug/l	92
43) Isopropyl Acetate	8.093	43	28512	3.63	ug/l #	85
95) Naphthalene	15.082	128	17700	2.23	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN041621\
Data File : VN066630.D
Acq On : 16 Apr 2021 16:56
Operator : JC/MD
Sample : PB135768TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135768TB

Quant Time: Apr 17 05:17:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041521W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 15 19:16:35 2021
Response via : Initial Calibration

